

(UK 2855)



V(0)a

United Kingdom of Great Britain and Northern Ireland

Certificate of EC type-approval of a measuring instrument

Number: UK 2855

issued by the Secretary of State for Innovation, Universities & Skills
Notified Body Number 0126

In accordance with the requirements of the Non-automatic Weighing Instruments Regulations 2000 (SI 2000/3236) which implement, in the United Kingdom, Council Directive 90/384/EEC, this certificate of EC type-approval has been issued to:

**Avery Weigh-Tronix Ltd
Foundry Lane
Smethwick
West Midlands B66 2LP
United Kingdom**

in respect of Class III non-automatic weighing instruments, designated the 350-Series, comprising the 350-Series indicator as described in Test Certificate GB-1303 and a weighing platform.

$n \leq 6000$ for Class III or IIII instruments

The necessary data (principal characteristics, alterations, securing, functioning etc) for identification purposes and conditions (when applicable) are set out in the descriptive annex to this certificate.

Signatory: P R Dixon
for Chief Executive
National Weights & Measures Laboratory
Department for Innovation, Universities & skills
Stanton Avenue
Teddington
Middlesex TW11 0JZ
United Kingdom

Date: 19 March 2009
Valid Until: 18 March 2019
Reference No: T1127/0026

Descriptive Annex

1 INTRODUCTION

This family of instruments utilises the digital indicating devices designated the GSE 350-Series indicators connected to a weighing platform to form single range, Class III or IIII, self-indicating, non-automatic weighing instruments (Figure 1).

2 FUNCTIONAL DESCRIPTION

2.1 Devices

The GSE 350-Series digital weight indicator is fully described in Test Certificate GB-1303.

2.2 Load cells

2.2.1 The following load cells may be used, in single-range applications only:

Model	Load cell E_{max}	Max capacity	$e \geq$	Max n
HBM PW2 C3	7.2 kg	5 kg	0.001 kg	3000
	12 kg	10 kg	0.002 kg	3000
	18 kg	16 kg	0.005 kg	3000
	36 kg	34 kg	0.01 kg	3000
	72 kg	70 kg	0.02 kg	3000
Vishay / Tede 1040 C3	5 kg	3 kg	0.001 kg	3000
	7 kg	5 kg	0.001 kg	3000
	10 kg	8 kg	0.001 kg	3000
	15 kg	13 kg	0.002 kg	3000
	20 kg	18 kg	0.002 kg	3000
	30 kg	28 kg	0.005 kg	3000
	50 kg	48 kg	0.005 kg	3000
	75 kg	73 kg	0.01 kg	3000
Vishay / Tede 1042 C4	5 kg	3 kg	0.0005 kg	4000
	7 kg	5 kg	0.001 kg	4000
	10 kg	8 kg	0.001 kg	4000
	15 kg	13 kg	0.002 kg	4000
Vishay / Tede 1042SYM C6	20 kg	18 kg	0.002 kg	6000
	30 kg	28 kg	0.005 kg	6000
	50 kg	48 kg	0.005 kg	6000
	75 kg	73 kg	0.01 kg	6000

2.2.2 Any compatible load cell(s) may be used providing the following conditions are met:

- There is a respective OIML Certificate of Conformity (R60) or a test certificate (EN45501) issued for the load cell by a Notified Body responsible for type examination under Directive 90/384/EEC.
- The certificate contains the load cell types and the necessary load cell data required for the manufacturer's declaration of compatibility of modules (WELMEC 2, Issue 4, 2004, No 11), and any particular installation requirements. A load cell marked NH is allowed only if humidity testing to EN45501 has been conducted on this loadcell.

- The compatibility of the load cells and indicator is established by the manufacturer by means of the compatibility of modules calculation, contained in the above WELMEC 2 document, at the time of verification or declaration of EC conformity of type.
- The loadcell transmission must conform to one of the examples shown in the WELMEC Guide 2.4, “Guide for Load cells”.

3 TECHNICAL DATA

3.1 Technical data for the indicator is provided in the Test Certificate (GB-1303).

4 PERIPHERAL DEVICES AND INTERFACES

4.1 Interfaces

The instrument has the following protected interfaces:

- RS232/RS485
- Ethernet (as a module connected to above port)
- Fibre optic module
- Digital inputs and outputs for interfacing with external equipment as follows:
 - 1 control input for initiating zero, print and/or tare from a remote switch
 - Additional 3-channel output module
 - 1 additional analogue output (0/4-20mA or 0-10V)

4.2 The weighing system may be connected to any peripheral device that has been issued with a test certificate by a Notified Body responsible for type approval under Directive 90/384/EEC in any Member State and bear the CE marking of conformity to the relevant directives; or

A peripheral device without a test certificate under the following conditions:

- it bears the CE marking for conformity to the EMC Directive 89/336/EEC;
- it is not capable of transmitting any data or instruction into the weighing instrument, other than to release a printout, checking for correct data transmission or validation;
- it prints weighing results and other data as received from the weighing instrument without any modification or further processing; and
- it complies with the applicable requirements of EN45501, i.e. 4.2, 4.4, 4.6 and 4.7.

5 APPROVAL CONDITIONS

The certificate is issued subject to the following conditions:

5.1 Legends and inscriptions

5.1.1 The instrument shall bear the following legends near the display of the weighing result:

Max
Min
e =
T (if ≠ - Max)

5.1.2 The instrument shall bear the following legends

- CE mark
- Green M
- Accuracy class
- Serial number
- Manufacturer's mark or name
- Certificate number

6 LOCATION OF SEALS AND VERIFICATION MARKS

6.1 The rating plate should be located on the indicator so that it is easily accessible and clearly visible in its regular operating position. The CE mark shall be impossible to remove without damaging it. The data plate shall be impossible to remove without it being destroyed.

The markings and inscriptions shall fulfil the requirements of Paragraph 1 of Annex IV of the Directive 90/384/EEC.

6.2 There are two methods of securing the instrument on verification, either by recording the audit trail counter or by setting the "PROG" switch on the main board to "NO" and applying a wire and seal as described in the technical manual. If the audit trail counter on the instrument increases above that recorded or if the recorded value is removed then this has to be considered as a broken seal. The two methods of securing are equivalent in the context of this approval.

6.3 Components that may not be dismantled or adjusted by the user (load cell connection, junction box) are secured by either a wire and seal, or by a tamper evident label and securing mark. The securing mark consists of a mark of the manufacturer and/or manufacturer's representative.

7 ALTERNATIVES

There are currently no alternatives.

8 CERTIFICATE HISTORY

ISSUE NO.	DATE	DESCRIPTION
UK 2855	19 March 2009	Type approval first issued.
-	-	No revisions have been issued.